



सेन्टर फॉर मेटिरियल्स फॉर इलेक्टॉनिक्स टेक्नोलाजी (सी-मेट)
CENTRE FOR MATERIALS FOR ELECTRONICS TECHNOLOGY (C-MET)
(Scientific Society, Ministry of Electronics and Information Technology (MeitY), Govt. of India)
IDA PHASE – III, CHERLAPALLY, HCL (PO), HYDERABAD – 500 051
Phone: +91-40-27265587, 27267006 Fax: +91-40-27261658 Web: www.cmet.gov.in

ENQUIRY LETTER

ENQ. NO.: HD/PUR/SP-40/CAB/03/2019-20/11245

Date: 30.08.2019

Subject: Procurement of 'Clean Air Bench' – Regarding

Due date/CLOSING for receipt of the quotations	23	09	2019	TIME BY:15.30 hours
DATE OF OPENING OF TECHNICAL BID	24	09	2019	TIME BY:12.00 hours

Sir,

Sealed quotations conforming to the specifications are invited for the following item(s). Quotation may please be submitted, keeping in view, the terms and conditions given below and overleaf, otherwise quotations will be rejected.

S. No.	Description of the item	Quantity	Specifications
01	Class 100 Wet Chemical Workbench	1 no.	As per 'Annexure – I' enclosed

Terms & Conditions:

C-MET is registered under Department of Scientific and Industrial Research (DSIR) and eligible for concessional GST rates @ 5% for Research Institutions as per Notification No. 45/2017- Central Tax (Rate) dated 14.11.2017 and Notification No. 10/2018-Integrated Tax (Rate) dated 25.01.2018. The requisite certificate for availing concessional GST Rate along with documents will be provided.

Note:

- The envelope super-scribed with [a] Quotation No. [b] Due Date & [c] Description of the Item and be sent to the undersigned before the due date.
- The Technical & Financial bids should be in separate sealed covers.**
- The quotation should be submitted in the enclosed format only.**
- The Enquiry Letter No. & date should be clearly indicated in the quotation, otherwise quotation will be rejected.**
- Firms or their authorized representatives are requested to attend the opening of **Combined/Technical bids** after closing time i.e. **on 24.09.2019 at 12:00 hours**. If there is any delay in opening of technical bids will be intimated to all bidders.
- These details are also available in our website: <http://www.cmet.gov.in> & CPPP Portal.

Thanking you,

Yours faithfully,

(G.K.VENKATESAN)
Administrative Officer
venkatesan@cmet.gov.in

Encl.: As above.

Specific terms and conditions to the tender:-

1. Please submit the Technical Details/Catalogue/Data Sheets.
 2. The offer should be valid for a period of 90 days from the date of opening of Tender.
 3. Please send the quotations ONLY in 'SEALED COVER' indicating our tender enquiry No. and due date by Hand/speed post/courier so as to reach us on or before the due date & time. C-MET will not be responsible for any postal delays.
 4. **E-Mail/fax quotations 'WILL NOT BE ACCEPTED'.**
 5. **All Tax invoices shall invariably bear your GST Registration No. (GSTIN) and the applicable GST rates. In the absence of which, the invoice(s) shall not be processed for payment.**
 6. Our standard delivery term is FOR, C-MET. In case any vendor offers delivery term of Ex-works, Packing and Forwarding charges if any should be indicated separately either as a percentage of the quoted rate or as a Lumpsum amount.
 7. **PAYMENT OF GST/IGST:** C-MET is registered under Department of Scientific and Industrial Research (DSIR) and eligible for concessional GST rates @ 5% for Research Institutions as per Notification No. 45/2017- Central Tax (Rate) AND 47/2017 - IT(Rate) both dated 14.11.2017. The requisite certificate for availing concessional IGST/GST Rate along with documents will be provided. **PAYMENT OF CUSTOM DUTY:** We are ELIGIBLE for concessional custom duty @ 5.15% as per DSIR issued certificate in terms of **Notification No. No. 51/96-Customs dated 23.07.1996** and necessary exemption certificate shall be issued upon request.
 8. Our standard payment term is 100% within 30 days for indigenous orders and sight draft for foreign orders. Firm should provide Bank details for RTGS remittance.
 9. **Liquidated Damages:** - The delivery period quoted should be realistic. The delivery period so quoted and mentioned in the order is the essence of the order/contract. In case of delay in delivery of material as per the delivery schedule, Liquidated Damage @ 0.5% per week or part thereof on the undelivered portion subject to a maximum of 10% of the contract value shall be levied. Wherever, installation and commissioning is also involved, the supply will be deemed to have been completed only when the entire stores is supplied, installed and accepted.
 10. **Security Deposit:-** Wherever the offer value is Rs. 5.00 lakhs or above, the successful tenderer should submit Security Deposit@10% of the order value by way of Bank Guarantee /FC Receipt. The Bank Guarantee shall be obtained from any Scheduled Bank on Rs. 200/- Non Judicial Stamp paper and should be valid beyond 2 months from the completion of all contractual obligations.
- Note: Clause NO. 9 & 10 are mandatory and offers of the vendors who have not agreed for the above conditions will be excluded from the procurement process. Micro and Small Vendors are not exempted from the submission of Security Deposit.**
11. In order to avail of the benefits extended to by Govt. of India to the Micro and Small Sectors, please submit attested copy of the valid Entrepreneur Memorandum Part-II signed by the General Manager, District Industries Centre/ Udyog Adhar / NSIC Registration Certification along with your offer.
 12. If any bidder submits forged / false document along with the tender, offer of such vendors will be summarily rejected and such bidders will be blacklisted for all future tenders.
 13. Wherever samples are required to be submitted along with the quotation, offer without sample will not be considered.
 14. **Price Reasonability Certificate:** Prescribed form enclosed.
 15. **Manufacturer's authorization letter:** A copy of **manufacturer's authorization letter** for participating in the bid should be attached.
- NOTE: Bidder should sign on each of the page as read and accepted thereof with signature and official stamp of the firm.**


Administrative Officer
C-MET, Hyderabad

CHECK LIST

Documents to be submitted along with the bids
(TECHNICAL BID)

S No	Description	Yes/No	Page No
1	Covering/introduction letter from the Firm referring to C-MET's tender enquiry No and date towards participation in the tender		
2	Acceptance of terms & conditions of tender enquiry		
3	Signed copy of tender enquiry		
4	Price Reasonability Certificate		
5	Bid form and Price schedule with commercial terms WITHOUT PRICE		
6	Manufacturer's Authorization letter		
7	Previous experience details alongwith documents like PO/WO etc. (not more than 3 years old)	NA	
8	Undertaking and declaration		
9	Technical compliance statement with/without deviation		
10	Technical brochure of the Product offered or drawings		
11	GST Registration certificate or No		
12	Bank details for NEFT/RTGS		
13	Warranty and additional warranty		

**Documents to be submitted alongwith Commercial bids
(PRICE BID)
(Separate sealed cover)**

S No	Description	Page No
1	Bid form and Price schedule with commercial terms WITH PRICE for main system	

**CENTRE FOR MATERIALS FOR ELECTRONICS TECHNOLOGY (C-MET)
HYDERABAD LABORATORY**

TECHNICAL SPECIFICATIONS FOR "CLASS 100 WET CHEMICAL WORKBENCH"

Description	Polypropylene Class 100 wet chemical workbench/station
Overall dimensions	1520 mm L X 1190 D x 2400 mm H
Construction	<ul style="list-style-type: none"> ➤ Material of construction is 10mm thick natural Polypropylene laminated sheets. ➤ Adjustable levelling feet's. ➤ An operator safety visor made of 6 mm thick clear acrylic sheet with PP frame, which is fixed in front of the bench and can be opened/closed based on the operator's requirement. ➤ Polypropylene drain of size 50mm with union connection at bottom of the sink base for draining ➤ Polypropylene C-Channels should be incorporated for improving strength & rigidity of wet bench. ➤ All materials selected for application in the most demanding of conditions and are suitable for a cleanroom application.
Work top for Acid usage	<p>Worktop should be made of 10mm thick polypropylene sheet with perforation. Appropriate supporting ribs should be given at the bottom of worktop to avoid deflection.</p> <p>Worktop spills, chemical spills and spent DI water should drain to the plenum below and from there to a dedicated main spigot to station rear.</p>
FFU(Fan Filter Unit)	<p>To provide class 100 clean air inside the wet bench, we intended to use imported Fan Filter module with HEPA filters. These FFM's are self-powered grid module with modular design, to fitting in standard T grid ceiling. Total height is 320mm. The Fan Filter Module is U.L listed and CE certified. FFU speed should vary from 0.2 to .5m/s and air flow rate vary from 460-760 CFM. The sound level should be 55dB's when measured from 760mm below from filter face. The vibration level should be 0.9mils.</p> <p>The Fan motor drive should be direct drive, forward curve centrifugal type with sealed bearing. The motor should have permanent split capacitor type, rated for continuous operation with thermal overload protection with two speed switch. The power requirement is 230V, 50Hz single phasewith maximum current of 1.9A with 280watts power input. The fan/motor assembly should be capable of delivering air at filter pressure of 9mm to 23mm final state. These FFM sshould have a speed controller for increasing the speed of the motor/blower from low, medium and high.</p> <p>Dimensions :1215 Lx 600 D x 350 mm H</p> <p>FFU Material of construction : Al Zinc Alloy</p> <p>HEPA filter should be rated 99.99% efficiency in removing 0.3μ or larger particles. Leak free in accordance with latest I.E.S recommended</p>

	<p>particle. The filter media is micro glass fibre with poly-string separator, sealed to casing. The filter guard should be provided with diamond pattern expanded sheet for protection.</p> <p>Dimensions :1210 Lx 600 D x 60 mm H</p> <p>Pre Filter rated 90% efficiency in removing 10 to 15 Micron particles. Pre filters should be made from non-woven materials.</p> <p>Dimensions :550 Lx 600 D x 50 mm H</p>
Gooseneck	1 nos. off Gooseneck made of Polypropylene with control valve for sink should be supplied.
Sink	Polypropylene sink should be supplied of dimensions 250 mm x250 mm x200 mm D
Exhaust grills	Exhaust grills should be provided at below the work top and above worktop. All the fumes generated at the work area should be exhausted through exhaust box.
Lighting	2 numbers of UV filtered lights/LED 2 x 36W should be supplied for adequate illumination, material of construction should be PVC/polycarbonate.
Bottom storage cabinet/chemical trolley	The station should be provided with 2 nos. of storage cabinet trollies for storing materials under the sink base.
Electrical	For circuit protection miniature circuit breakers (MCB) will be provided. There should be 4 nos. of 15amps socket and switches provided on front of the station on either side.
Magnehelic gauge	Magnehelic gauge of 50mm of water column capacity to measure the differential pressure at the exhaust plenum box in fume hood.
Exhaust port	Unique exhaust port design ensures that the fumes should be exhausted smoothly without any turbulence at the exhaust port. Also it must ensure low noise level.
Exhaust Blower	SISW FRP Centrifugal Blower @ 900 CFM @ 75 mm Aq complete with 3 ph TEFC Motor & all Std. accessory & mountings
PVC Ducting	3 mm thick 200 mm dia. duct 12 meters should be supplied with required fittings.
Optional	Teflon N2 Gun 1 nos., DI Water gun
Installation, Testing, Commissioning & validation.	All the equipment's shall be installed, tested, commissioned and validated at C-MET premises.
Warranty	One year after installation and commissioning OR 18 months after delivery.

(TO BE SUBMITTED ALONG WITH TECHNICAL BID)

(in the company letter head)

Technical compliance statement in r/o CLASS 100 WET CHEMICAL WORKBENCH

Technical Requirement	Technical Specification	Specifications of quoted Model	Deviation, if any, to be indicated in unambiguous terms	Remarks
Description	Polypropylene Class 100 wet chemical workbench/station			
Overall dimensions	1520 mm L X 1190 D x 2400 mm H			
Construction	<ul style="list-style-type: none"> ➤ Material of construction is 10mm thick natural Polypropylene laminated sheets. ➤ Adjustable levelling feet's. ➤ An operator safety visor made of 6 mm thick clear acrylic sheet with PP frame, which is fixed in front of the bench and can be opened/closed based on the operator's requirement. ➤ Polypropylene drain of size 50mm with union connection at bottom of the sink base for draining ➤ Polypropylene C-Channels should be incorporated for improving strength & rigidity of wet bench. ➤ All materials selected for application in the most demanding of conditions and are suitable for a cleanroom application. 			

Work top for Acid usage	Worktop should be made of 10mm thick polypropylene sheet with perforation. Appropriate supporting ribs should be given at the bottom of worktop to avoid deflection. Worktop spills, chemical spills and spent DI water should drain to the plenum below and from there to a dedicated main spigot to station rear.		
FFU(Fan Filter Unit)	<p>To provide class 100 clean air inside the wet bench, we intended to use imported Fan Filter module with HEPA filters. These FFM's are self-powered grid module with modular design, to fitting in standard T grid ceiling. Total height is 320mm. The Fan Filter Module is U.L listed and CE certified. FFU speed should vary from 0.2 to .5m/s and air flow rate vary from 460-760 CFM. The sound level should be 55dB's when measured from 760mm below from filter face. The vibration level should be 0.9mils.</p> <p>The Fan motor drive should be direct drive, forward curve centrifugal type with sealed bearing. The motor should have permanent split capacitor type, rated for continuous operation with thermal overload protection with two speed switch. The power requirement is 230V, 50Hz single phase with maximum current of 1.9A with 280watts power input.</p> <p>The fan/motor assembly should be capable of delivering air at filter pressure of 9mm to 23mm final state. These FFM should have a speed controller for increasing the speed of the motor/blower from low, medium and high.</p>		

	<p>Dimensions :1215 Lx 600 D x 350 mm H</p> <p>FFU Material of construction : Al Zinc Alloy</p> <p>HEPA filter should be rated 99.99% efficiency in removing 0.3µ or larger particles. Leak free in accordance with latest I.E.S recommended particle. The filter media is micro glass fibre with poly-string separator, sealed to casing. The filter guard should be provided with diamond pattern expanded sheet for protection.</p> <p>Dimensions :1210 Lx 600 D x 60 mm H</p> <p>Pre Filter rated 90% efficiency in removing 10 to 15 Micron particles. Pre filters should be made from non-woven materials.</p> <p>Dimensions :550 Lx 600 D x 50 mm H</p>			
Gooseneck	1 nos. off Gooseneck made of Polypropylene with control valve for sink should be supplied.			
Sink	Polypropylene sink should be supplied of dimensions 250 mm x250 mm x200 mm D			
Exhaust grills	Exhaust grills should be provided at below the work top and above worktop. All the fumes generated at the work area should be exhausted through exhaust box.			
Lighting	2 numbers of UV filtered lights/LED 2 x 36W should be supplied for adequate illumination, material of construction should be PVC/polycarbonate.			
Bottom storage cabinet/chemical trolley	The station should be provided with 2 nos. of storage cabinet trollies for storing materials under the sink base.			
Electrical	For circuit protection miniature circuit breakers (MCB) will be provided. There should be 4 nos. of 15amps socket and switches provided on front of the station on			

	either side.			
Magnehelic gauge	Magnehelic gauge of 50mm of water column capacity to measure the differential pressure at the exhaust plenum box in fume hood.			
Exhaust port	Unique exhaust port design ensures that the fumes should be exhausted smoothly without any turbulence at the exhaust port. Also it must ensures low noise level.			
Exhaust Blower	SISW FRP Centrifugal Blower @ 900 CFM @ 75 mm Aq complete with 3 ph TEFC Motor & all Std. accessory& mountings			
PVC Ducting	3 mm thick 200 mm dia. duct 12 meters should be supplied with required fittings.			
Optional	Teflon N2 Gun 1 nos., DI Water gun			
Installation, Testing, Commissioning & validation.	All the equipment's shall be installed, tested, commissioned and validated at C-MET premises.			
Warranty	One year after installation and commissioning OR 18 months after delivery			

Signature of Tenderer _____

Name _____

Seal of the tenderer _____

(in the company letter head)

BID FORM AND PRICE SCHEDULE

Tender Reference No. : ENQ. NO.: HD/PUR/SP-40/CAB/05/2019-20/11245
Date : 30.08.2019

Name of the Stores/
Goods/Equipment : Clean Air Bench

Sl. No.	Description	Qty	HSN CODE	Rate/Rs.	Amount/Rs.
1	Class 100 Wet Chemical Workbench	1 no.			
	GST				
	Freight Charges				
	Packing & Forwarding				
	Transit Insurance				
	TOTAL				

1. Payment Terms:
2. Bid Validity:
3. Delivery Schedule:
4. Place of delivery:
5. Warranty, if applicable:
6. GST Registration No.:
7. Any other relevant information:

Signature of Tenderer _____

Name _____

Seal of the tenderer _____

(in the company letter head)

BIDDER'S DETAILS

1	Name of the supplied/party/firm	
2	Name of the authorized representative	
3	Communication address	
4	Telephone Nos	
5	Mobile Nos	
6	Fax and e-mail address	
7	Web address	
8	Bank details for payment through NEFT/RTGS	
9	Name of Bank	
	Branch	
	Account No	
	IFSC No	
	MICR No	
	Cancelled cheque for verification of above details	

(Signature & seal of the bidder)

(in the company letter head)

PRICE REASONABILITY CERTIFICATE

It is certified that the rates quoted _____ are not more than as charged to other Govt. /PSU's for similar supplies made in recent past. Copies of the supply orders received from other Govt/PSU's are enclosed alongwith the technical bid.

Signature of Tenderer _____

Name _____

Business Address _____

Seal of the Tenderer _____

(TO BE SUBMITTED ALONG WITH TECHNICAL BID)

(in the company letter head)

To,

Date:

The Administrative Officer
C-MET, IDA Phase III
Cherlapally,
Hyderabad 500 051

Sub: Acceptance of Terms & Conditions of Tender.

Tender Reference No. _____ dated _____

Name of the Tender/Work:

Sir/Madam,

I/We have downloaded/obtained the tender document(s) for the above mentioned 'Tender/Work' for the web site(s) namely:

as per your advertisement, given in the mentioned website(s)/enquiry letter.

2. I/we hereby certify that I/we have read the entire terms & conditions of the tender documents from Page No. _____ to _____ (including all documents like Annexure(s), schedules(s), etc.) which form part of the contract agreement and I/We shall abide hereby by the terms/conditions/clauses contained therein.

3. The corrigendum(s) issued from time to time by your department/organization too have all been taken into consideration, while submitting this acceptance letter.

4. I/We hereby unconditionally accept the tender conditions of above mentioned tender document (s)/Corrigendum(s) in its totally/entirely.

5. In case any provisos of this tender are found violated, then your Department/Organizations shall without prejudice to any other right or remedy be at liberty to reject this tender/bid including the forfeiture of the full said earnest money deposit absolutely.

Yours faithfully,

Name & Signature of the Supplier/Contractor/
Authorized signatory with Official stamp/seal

(TO BE SUBMITTED ALONG WITH TECHNICAL BID)

(in the company letter head)

UNDERTAKING/ DECLARATION

I/We having our office at _____ declare that I/we have never been blacklisted by any State Government/Central Government or any State/Central PSU.

2. I/We, _____ hereby declare that the particulars furnished by me/us in this offer are true to the best of my/our knowledge and I/We understand and accept that, if at any stage the information furnished by me/us are found to be incorrect or false, I/We am/are liable for disqualification from this tender and also liable for any penal action that may arise due to the above, besides being black listed.

3. I/We _____ hereby accept to provide Warranty/Guarantee, additional Warranty, Comprehensive AMC

4. This is also certified that the rates quoted in the financial bid are not more than the rates charged from any other Institution/Department/ Organization in the last six months. No Other charges would be payable by the Institute.

5. I/We _____ have gone through the terms and conditions and will abide by them as laid down above. I hereby undertake that the information provided above and elsewhere in the tender is true and the tender is liable to rejection if the same is found to be false or the information is found to have been suppressed by me.

Name & Signature of the Supplier/Contractor/
Authorized signatory with Official stamp/seal

**(NOTE: WITHOUT THIS BID WILL BE REJECT - TO BE SUBMITTED
ALONGWITH TECHNICAL BID)
(in the company letter head)**

MANUFACTURER'S AUTHORISATION FORM

The Administrative Officer
C-MET, IDA Phase-III
Cherlapally, HCL Post
Hyderabad – 500 051
TELANGANA STATE,
India

Dear Sir,

Ref. Your TE document No _____, dated _____

We, _____ who are proven and reputable manufacturers of _____ (*name and description of the goods offered in the tender*) having factories at _____, hereby authorize Messrs _____ (*name and address of the agent*) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We further confirm that no supplier or firm or individual other than Messrs. _____ (*name and address of the above agent*) is authorized to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also hereby extend our full warranty as applicable as per tender clause for the goods and services offered for supply by the above firm against this TE document.

We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorized agent.

We also confirm that the price quoted by our agent shall not exceed than that which we would have quoted directly.

Yours faithfully,

[Signature with date, name and designation]
For and on behalf of Messrs _____

[Name & address of the manufacturers]

Note: 1. This letter of authorization should be on the letterhead of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
2. Original letter may be sent.